Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE				
Calibration File No.:	00833	Calib. Date:	10/10/2014	Calib. No.:	00017
Certification File No.:	00834	Cert. Date:	10/10/2014	Cert. No.:	00014
Linearity File No.:	00835	Lin. Date:	10/10/2014	Lin. No.:	00014
Solution File No.:	00877	Soln. Date:	03/01/2015	Soln. No.:	00172
Sequential File No.:	00877	File Date:	03/01/2015		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAB-0015
Control Solution %:	0.100%			Expires:	04/29/2016
Solution Control Lot:	14D129			Bottle No.:	0133

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:13S		
Control 1 EC	0.101%	10:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15S		
Control 2 EC	0.100%	10:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16S		
Control 3 EC	0.100%	10:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRADY First Name: PATRICK MI: R

Signature: PR #175 Badge No.: 175

Date: 03/01/2015

PR #175