

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01191	Calib. No.: 00024
Certification File No.:	01192	Cert. No.: 00016
Linearity File No.:	01193	Lin. No.: 00016
Solution File No.:	01257	Soln. No.: 00208
Sequential File No.:	01257	File Date: 03/19/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	14D129	Expires: 04/29/2016
		Bottle No.: 0143

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:37S		
Control 1 EC	0.098%	20:38S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:39S		
Control 2 EC	0.099%	20:39S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:40S		
Control 3 EC	0.099%	20:41S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 03/19/2015