

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 01259 Calib. Date: 03/25/2015 Calib. No.: 00025  
Certification File No.: 01260 Cert. Date: 03/25/2015 Cert. No.: 00017  
Linearity File No.: 01261 Lin. Date: 03/25/2015 Lin. No.: 00017  
Solution File No.: 01280 Soln. Date: 05/21/2015 Soln. No.: 00211  
Sequential File No.: 01280 File Date: 05/21/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113  
Control Solution %: 0.100% Expires: 04/29/2016  
Solution Control Lot: 14D129 Bottle No.: 0144

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:57D		
Control 1 EC	0.100%	11:58D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:58D		
Control 2 EC	0.099%	11:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59D		
Control 3 EC	0.099%	12:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:01D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PHK #129*

Badge No.: 129

Date: 05/21/2015