

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00886	Calib. Date: 03/25/2015
Certification File No.:	00887	Calib. No.: 00018
Linearity File No.:	00888	Cert. Date: 03/25/2015
Solution File No.:	00900	Cert. No.: 00015
Sequential File No.:	00900	Lin. Date: 03/25/2015
		Lin. No.: 00015
		Soln. Date: 05/21/2015
		Soln. No.: 00176
		File Date: 05/21/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	14D129	Expires: 04/29/2016
		Bottle No.: 0130

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:58D		
Control 1 EC	0.100%	11:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59D		
Control 2 EC	0.099%	12:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:01D		
Control 3 EC	0.100%	12:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_



Badge No.: 129

Date: 05/21/2015