

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01259 Calib. Date: 03/25/2015 Calib. No.: 00025
Certification File No.: 01260 Cert. Date: 03/25/2015 Cert. No.: 00017
Linearity File No.: 01261 Lin. Date: 03/25/2015 Lin. No.: 00017
Solution File No.: 01283 Soln. Date: 05/22/2015 Soln. No.: 00212
Sequential File No.: 01283 File Date: 05/22/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 08/22/2016
Solution Control Lot: 14H132 Bottle No.: 0056

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:54D		
Control 1 EC	0.099%	20:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:56D		
Control 2 EC	0.099%	20:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:57D		
Control 3 EC	0.098%	20:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: 

Badge No.: 178

Date: 05/22/2015