

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00886 Calib. Date: 03/25/2015 Calib. No.: 00018
Certification File No.: 00887 Cert. Date: 03/25/2015 Cert. No.: 00015
Linearity File No.: 00888 Lin. Date: 03/25/2015 Lin. No.: 00015
Solution File No.: 00901 Soln. Date: 05/22/2015 Soln. No.: 00177
Sequential File No.: 00901 File Date: 05/22/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 08/22/2016
Solution Control Lot: 14H132 Bottle No.: 0051

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 20:38D | | |
| Control 1 EC | 0.101% | 20:39D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 20:39D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:40D | | |
| Control 2 EC | 0.100% | 20:40D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 20:40D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:41D | | |
| Control 3 EC | 0.100% | 20:42D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 20:42D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 20:42D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: 

Badge No.: 178

Date: 05/22/2015