

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00886 Calib. Date: 03/25/2015 Calib. No.: 00018
Certification File No.: 00887 Cert. Date: 03/25/2015 Cert. No.: 00015
Linearity File No.: 00888 Lin. Date: 03/25/2015 Lin. No.: 00015
Solution File No.: 00930 Soln. Date: 07/03/2015 Soln. No.: 00180
Sequential File No.: 00930 File Date: 07/03/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 08/22/2016
Solution Control Lot: 14H132 Bottle No.: 0067

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:45D		
Control 1 EC	0.100%	11:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46D		
Control 2 EC	0.099%	11:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48D		
Control 3 EC	0.099%	11:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 07/03/2015