

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01259	Calib. Date:	03/25/2015
Certification File No.:	01260	Cert. Date:	03/25/2015
Linearity File No.:	01261	Lin. Date:	03/25/2015
Solution File No.:	01327	Soln. Date:	07/10/2015
Sequential File No.:	01327	File Date:	07/10/2015

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXB S3-113
Control Solution %:	0.100%			Expires:	08/22/2016
Solution Control Lot:	14H132			Bottle No.:	0072

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:27D		
Control 1 EC	0.099%	13:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:28D		
Control 2 EC	0.099%	13:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:29D		
Control 3 EC	0.099%	13:30D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:31D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: 

Badge No.: 129

Date: 07/10/2015