

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00886	Calib. Date: 03/25/2015
Certification File No.:	00887	Calib. No.: 00018
Linearity File No.:	00888	Cert. Date: 03/25/2015
Solution File No.:	00940	Cert. No.: 00015
Sequential File No.:	00940	Lin. Date: 03/25/2015
		Lin. No.: 00015
		Soln. Date: 07/10/2015
		Soln. No.: 00181
		File Date: 07/10/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	14H132	Expires: 08/22/2016
		Bottle No.: 0073

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:30D		
Control 1 EC	0.102%	13:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:32D		
Control 2 EC	0.101%	13:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:33D		
Control 3 EC	0.101%	13:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

PK #129

Badge No.: 129

Date: 07/10/2015