

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00886	Calib. Date:	03/25/2015
Certification File No.:	00887	Cert. Date:	03/25/2015
Linearity File No.:	00888	Lin. Date:	03/25/2015
Solution File No.:	00955	Soln. Date:	07/23/2015
Sequential File No.:	00955	File Date:	07/23/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	14H132	Expires:	08/22/2016
		Bottle No.:	0063

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:54D		
Control 1 EC	0.101%	14:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55D		
Control 2 EC	0.100%	14:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57D		
Control 3 EC	0.100%	14:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____



Badge No.: 129

Date: 07/23/2015