

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01259	Calib. Date:	03/25/2015
Certification File No.:	01260	Cert. No.:	00025
Linearity File No.:	01261	Cert. Date:	03/25/2015
Solution File No.:	01344	Lin. Date:	03/25/2015
Sequential File No.:	01344	Lin. No.:	00017
		Soln. Date:	08/07/2015
		Soln. No.:	00219
		File Date:	08/07/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	14H132	Expires:	08/22/2016
		Bottle No.:	0053

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:23D		
Control 1 EC	0.100%	14:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:24D		
Control 2 EC	0.100%	14:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:26D		
Control 3 EC	0.099%	14:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

Badge No.: 129

Date: 08/07/2015