

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01259	Calib. Date: 03/25/2015
Certification File No.:	01260	Calib. No.: 00025
Linearity File No.:	01261	Cert. Date: 03/25/2015
Solution File No.:	01351	Cert. No.: 00017
Sequential File No.:	01351	Lin. Date: 03/25/2015
		Lin. No.: 00017
		Soln. Date: 08/20/2015
		Soln. No.: 00220
		File Date: 08/20/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	14H132	Expires: 08/22/2016
		Bottle No.: 0059

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:10D		
Control 1 EC	0.100%	07:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:11D		
Control 2 EC	0.099%	07:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:13D		
Control 3 EC	0.099%	07:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

Badge No.: 129

Date: 08/20/2015