

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
<b>Location:</b>	WALL TOWNSHIP POLICE	
<b>Calibration File No.:</b>	00886	<b>Calib. Date:</b> 03/25/2015
<b>Certification File No.:</b>	00887	<b>Calib. No.:</b> 00018
<b>Linearity File No.:</b>	00888	<b>Cert. Date:</b> 03/25/2015
<b>Solution File No.:</b>	00980	<b>Cert. No.:</b> 00015
<b>Sequential File No.:</b>	00980	<b>Lin. Date:</b> 03/25/2015
		<b>Lin. No.:</b> 00015
		<b>Soln. Date:</b> 08/20/2015
		<b>Soln. No.:</b> 00184
		<b>File Date:</b> 08/20/2015

<b>Calibrating Unit:</b>	WET	<b>Model No.:</b> CU-34	<b>Serial No.:</b> DDXB S3-114
<b>Control Solution %:</b>	0.100%		<b>Expires:</b> 08/22/2016
<b>Solution Control Lot:</b>	14H132		<b>Bottle No.:</b> 0065

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:14D		
Control 1 EC	0.101%	07:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:15D		
Control 2 EC	0.100%	07:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:16D		
Control 3 EC	0.100%	07:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: \_\_\_\_\_

*PTL K #129*

Badge No.: 129

Date: 08/20/2015