

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0037 |
| Location: | WALL TOWNSHIP POLICE | |
| Calibration File No.: | 01259 | Calib. Date: 03/25/2015 |
| Certification File No.: | 01260 | Calib. No.: 00025 |
| Linearity File No.: | 01261 | Cert. Date: 03/25/2015 |
| Solution File No.: | 01362 | Cert. No.: 00017 |
| Sequential File No.: | 01362 | Lin. Date: 03/25/2015 |
| | | Lin. No.: 00017 |
| | | Soln. Date: 09/06/2015 |
| | | Soln. No.: 00222 |
| | | File Date: 09/06/2015 |

| | | | |
|-----------------------|--------|------------------|-------------------------|
| Calibrating Unit: | WET | Model No.: CU-34 | Serial No.: DDXB S3-113 |
| Control Solution %: | 0.100% | | Expires: 06/13/2017 |
| Solution Control Lot: | 15F139 | | Bottle No.: 0479 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 04:53D | | |
| Control 1 EC | 0.097% | 04:54D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 04:54D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:54D | | |
| Control 2 EC | 0.096% | 04:55D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 04:55D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:56D | | |
| Control 3 EC | 0.096% | 04:56D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 04:56D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:57D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: _____



Badge No.: 174

Date: 09/06/2015