

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00995	Calib. Date:	09/08/2015
Certification File No.:	00996	Cert. Date:	09/08/2015
Linearity File No.:	00997	Lin. Date:	09/08/2015
Solution File No.:	00998	Soln. Date:	09/08/2015
Sequential File No.:	00998	File Date:	09/08/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	14H132	Expires:	08/22/2016
		Bottle No.:	0054

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:14D		
Control 1 EC	0.101%	14:15D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:15D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:16D		
Control 2 EC	0.101%	14:16D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:16D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:17D		
Control 3 EC	0.101%	14:18D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:18D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXK P2-393 ⁽²⁾

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: _____

SGT. P. Dennis #5925

Badge No.: 5925

Date: 09/08/2015