

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00995	Calib. Date: 09/08/2015
Certification File No.:	00996	Calib. No.: 00019
Linearity File No.:	00997	Cert. Date: 09/08/2015
Solution File No.:	01009	Cert. No.: 00016
Sequential File No.:	01009	Lin. Date: 09/08/2015
		Lin. No.: 00016
		Soln. Date: 09/20/2015
		Soln. No.: 00187
		File Date: 09/20/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	14H132	Expires: 08/22/2016
		Bottle No.: 0060

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:06D		
Control 1 EC	0.100%	09:07D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:07D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:07D		
Control 2 EC	0.100%	09:08D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:09D		
Control 3 EC	0.100%	09:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: _____

Sgt Daniel Santoro
165

Badge No.: 165

Date: 09/20/2015

Emilio