

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00995	Calib. Date: 09/08/2015
Certification File No.:	00996	Calib. No.: 00019
Linearity File No.:	00997	Cert. Date: 09/08/2015
Solution File No.:	01011	Cert. No.: 00016
Sequential File No.:	01011	Lin. Date: 09/08/2015
		Lin. No.: 00016
		Soln. Date: 09/25/2015
		Soln. No.: 00188
		File Date: 09/25/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	14H132	Expires: 08/22/2016
		Bottle No.: 0075

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:07D		
Control 1 EC	0.101%	11:08D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:08D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:08D		
Control 2 EC	0.100%	11:09D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:09D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:10D		
Control 3 EC	0.099%	11:10D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:10D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 09/25/2015