

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00995	Calib. Date: 09/08/2015
Certification File No.:	00996	Calib. No.: 00019
Linearity File No.:	00997	Cert. Date: 09/08/2015
Solution File No.:	01022	Cert. No.: 00016
Sequential File No.:	01022	Lin. Date: 09/08/2015
		Lin. No.: 00016
		Soln. Date: 10/09/2015
		Soln. No.: 00189
		File Date: 10/09/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	14H132	Expires: 08/22/2016
		Bottle No.: 0064

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:53D		
Control 1 EC	0.100%	09:54D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:54D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:55D		
Control 2 EC	0.099%	09:55D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:55D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56D		
Control 3 EC	0.100%	09:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: \_\_\_\_\_



Badge No.: 129

Date: 10/09/2015