

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01363	Calib. Date: 09/08/2015
Certification File No.:	01364	Calib. No.: 00026
Linearity File No.:	01365	Cert. Date: 09/08/2015
Solution File No.:	01377	Cert. No.: 00018
Sequential File No.:	01377	Lin. Date: 09/08/2015
		Lin. No.: 00018
		Soln. Date: 10/16/2015
		Soln. No.: 00226
		File Date: 10/16/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	15F139	Expires: 06/13/2017
		Bottle No.: 0480

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:53D		
Control 1 EC	0.097%	21:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:54D		
Control 2 EC	0.095%	21:55D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:55D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:56D		
Control 3 EC	0.096%	21:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	21:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: REINECKE

First Name: VINCENT

MI: F

Signature: _____

Badge No.: 180

Date: 10/16/2015