

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00995	Calib. Date:	09/08/2015
Certification File No.:	00996	Cert. Date:	09/08/2015
Linearity File No.:	00997	Lin. Date:	09/08/2015
Solution File No.:	01035	Soln. Date:	10/29/2015
Sequential File No.:	01035	File Date:	10/29/2015

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAB-0015
Control Solution %:	0.100%			Expires:	08/22/2016
Solution Control Lot:	14H132			Bottle No.:	0070

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:33S		
Control 1 EC	0.100%	13:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:35S		
Control 2 EC	0.100%	13:35S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:35S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:36S		
Control 3 EC	0.100%	13:37S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:37S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

PTL. K #129

Badge No.: 129

Date: 10/29/2015