

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00995	Calib. Date:	09/08/2015
Certification File No.:	00996	Cert. Date:	09/08/2015
Linearity File No.:	00997	Lin. Date:	09/08/2015
Solution File No.:	01041	Soln. Date:	11/20/2015
Sequential File No.:	01041	File Date:	11/20/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	15F139	Expires:	06/13/2017
		Bottle No.:	1354

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:27S		
Control 1 EC	0.098%	16:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	16:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:28S		
Control 2 EC	0.097%	16:29S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	16:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:29S		
Control 3 EC	0.096%	16:30S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.097%	16:30S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____



Badge No.: 129

Date: 11/20/2015