

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00995	Calib. Date: 09/08/2015
Certification File No.:	00996	Calib. No.: 00019
Linearity File No.:	00997	Cert. Date: 09/08/2015
Solution File No.:	01044	Cert. No.: 00016
Sequential File No.:	01044	Lin. Date: 09/08/2015
		Lin. No.: 00016
		Soln. Date: 11/26/2015
		Soln. No.: 00194
		File Date: 11/26/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDAB-0015
Control Solution %:	0.100%		Expires: 06/13/2017
Solution Control Lot:	15F139		Bottle No.: 1360

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:20S		
Control 1 EC	0.098%	01:20S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:21S		
Control 2 EC	0.096%	01:22S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	01:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:22S		
Control 3 EC	0.096%	01:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	01:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: 

Badge No.: 177

Date: 11/26/2015