

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01363	Calib. Date: 09/08/2015
Certification File No.:	01364	Calib. No.: 00026
Linearity File No.:	01365	Cert. Date: 09/08/2015
Solution File No.:	01395	Cert. No.: 00018
Sequential File No.:	01395	Lin. Date: 09/08/2015
		Lin. No.: 00018
		Soln. Date: 11/29/2015
		Soln. No.: 00231
		File Date: 11/29/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	15F139	Expires: 06/13/2017
		Bottle No.: 1352

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:56S		
Control 1 EC	0.097%	04:57S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:57S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:57S		
Control 2 EC	0.097%	04:58S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:58S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:59S		
Control 3 EC	0.097%	04:59S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.097%	04:59S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: _____

Sgt Daniel Santoro

Badge No.: 165

Date: 11/29/2015