

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01363	Calib. Date: 09/08/2015
Certification File No.:	01364	Calib. No.: 00026
Linearity File No.:	01365	Cert. Date: 09/08/2015
Solution File No.:	01401	Cert. No.: 00018
Sequential File No.:	01401	Lin. Date: 09/08/2015
		Lin. No.: 00018
		Soln. Date: 12/18/2015
		Soln. No.: 00232
		File Date: 12/18/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	15F139	Expires: 06/13/2017
		Bottle No.: 1351

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:37S		
Control 1 EC	0.097%	10:38S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:38S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39S		
Control 2 EC	0.097%	10:39S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:40S		
Control 3 EC	0.097%	10:41S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL.

MI: A.

Signature: \_\_\_\_\_

*PTL K #129*

Badge No.: 129

Date: 12/18/2015