

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01363	Calib. Date: 09/08/2015
Certification File No.:	01364	Cert. No.: 00026
Linearity File No.:	01365	Cert. Date: 09/08/2015
Solution File No.:	01403	Lin. No.: 00018
Sequential File No.:	01403	Lin. Date: 09/08/2015
		Soln. No.: 00233
		Soln. Date: 12/21/2015
		File Date: 12/21/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	15J149	Expires: 10/01/2017
		Bottle No.: 0953

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:10S		
Control 1 EC	0.100%	01:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:11S		
Control 2 EC	0.099%	01:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	01:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:13S		
Control 3 EC	0.099%	01:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	01:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BRADY

First Name: PATRICK

MI: R

Signature:           *PRC* #175          

Badge No.: 175

Date: 12/21/2015