

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00995	Calib. Date:	09/08/2015
Certification File No.:	00996	Cert. Date:	09/08/2015
Linearity File No.:	00997	Lin. Date:	09/08/2015
Solution File No.:	01060	Soln. Date:	12/22/2015
Sequential File No.:	01060	File Date:	12/22/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	15J149	Expires:	10/01/2017
		Bottle No.:	0952

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:58S		
Control 1 EC	0.100%	22:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:59S		
Control 2 EC	0.098%	23:00S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:00S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:01S		
Control 3 EC	0.099%	23:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: KELLY

First Name: JASON

MI: R

Signature: \_\_\_\_\_



Badge No.: 166

Date: 12/22/2015