

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0037 |
| Location: | WALL TOWNSHIP POLICE | |
| Calibration File No.: | 01363 | Calib. Date: 09/08/2015 |
| Certification File No.: | 01364 | Calib. No.: 00026 |
| Linearity File No.: | 01365 | Cert. Date: 09/08/2015 |
| Solution File No.: | 01408 | Cert. No.: 00018 |
| Sequential File No.: | 01408 | Lin. Date: 09/08/2015 |
| | | Lin. No.: 00018 |
| | | Soln. Date: 12/25/2015 |
| | | Soln. No.: 00234 |
| | | File Date: 12/25/2015 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDAB-0014 |
| Solution Control Lot: | 15J149 | Expires: 10/01/2017 |
| | | Bottle No.: 0960 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 01:43S | | |
| Control 1 EC | 0.100% | 01:44S | 34.1°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 01:44S | 34.1°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:44S | | |
| Control 2 EC | 0.099% | 01:45S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 01:45S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:46S | | |
| Control 3 EC | 0.099% | 01:46S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 01:46S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 01:47S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 12/25/2015