

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
<b>Location:</b>	WALL TOWNSHIP POLICE	
Calibration File No.:	00995	Calib. Date: 09/08/2015
Certification File No.:	00996	Cert. No.: 00019
Linearity File No.:	00997	Cert. Date: 09/08/2015
Solution File No.:	01066	Lin. No.: 00016
Sequential File No.:	01066	Lin. Date: 09/08/2015
		Soln. No.: 00201
		Soln. Date: 12/28/2015
		File Date: 12/28/2015
<b>Calibrating Unit:</b>	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	15J149	Expires: 10/01/2017
		Bottle No.: 0956

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:11S		
Control 1 EC	0.099%	04:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:12S		
Control 2 EC	0.099%	04:13S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:13S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:14S		
Control 3 EC	0.099%	04:14S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:14S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: \_\_\_\_\_



Badge No.: 174

Date: 12/28/2015