

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00995	Calib. Date: 09/08/2015
Certification File No.:	00996	Calib. No.: 00019
Linearity File No.:	00997	Cert. Date: 09/08/2015
Solution File No.:	01097	Cert. No.: 00016
Sequential File No.:	01097	Lin. Date: 09/08/2015
		Lin. No.: 00016
		Soln. Date: 02/01/2016
		Soln. No.: 00207
		File Date: 02/01/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAB-0015
Control Solution %:	0.100%			Expires:	12/04/2016
Solution Control Lot:	14L134			Bottle No.:	1173

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:46S		
Control 1 EC	0.100%	23:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:48S		
Control 2 EC	0.100%	23:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:49S		
Control 3 EC	0.100%	23:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GRAMLICH

First Name: JONATHAN

MI: R

Signature: _____



Badge No.: 179

Date: 02/01/2016