

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01363	Calib. Date:	09/08/2015
Certification File No.:	01364	Cert. Date:	09/08/2015
Linearity File No.:	01365	Lin. Date:	09/08/2015
Solution File No.:	01412	Soln. Date:	01/01/2016
Sequential File No.:	01412	File Date:	01/01/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	15J149	Expires:	10/01/2017
		Bottle No.:	9630

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:10S		
Control 1 EC	0.101%	09:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:11S		
Control 2 EC	0.100%	09:12S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:12S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13S		
Control 3 EC	0.099%	09:13S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:13S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MALONE

First Name: MICHAEL

MI: T

Signature: *Michael Malone FIS*

Badge No.: 153

Date: 01/01/2016