

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01363	Calib. Date:	09/08/2015
Certification File No.:	01364	Cert. Date:	09/08/2015
Linearity File No.:	01365	Lin. Date:	09/08/2015
Solution File No.:	01417	Soln. Date:	01/04/2016
Sequential File No.:	01417	File Date:	01/04/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	15J149	Expires:	10/01/2017
		Bottle No.:	0965

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:06S		
Control 1 EC	0.100%	16:06S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:06S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:07S		
Control 2 EC	0.099%	16:08S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:08S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:08S		
Control 3 EC	0.099%	16:09S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:09S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KLIMIK

First Name: DAVID

MI: W

Signature: \_\_\_\_\_

Tom # D. Klimik #7040

Badge No.: 7040

Date: 01/04/2016