

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00995	Calib. Date:	09/08/2015
Certification File No.:	00996	Cert. Date:	09/08/2015
Linearity File No.:	00997	Lin. Date:	09/08/2015
Solution File No.:	01073	Soln. Date:	01/04/2016
Sequential File No.:	01073	File Date:	01/04/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	14B127	Expires:	02/14/2016
		Bottle No.:	0745

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:37S		
Control 1 EC	0.101%	16:38S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:38S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:38S		
Control 2 EC	0.100%	16:39S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:40S		
Control 3 EC	0.100%	16:41S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:41S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KLIMIK

First Name: DAVID

MI: W

Signature: _____

TPA II D. Klimik #7040

Badge No.: 7040

Date: 01/04/2016