

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
**Location:** WALL TOWNSHIP POLICE  
**Calibration File No.:** 01363 **Calib. Date:** 09/08/2015 **Calib. No.:** 00026  
**Certification File No.:** 01364 **Cert. Date:** 09/08/2015 **Cert. No.:** 00018  
**Linearity File No.:** 01365 **Lin. Date:** 09/08/2015 **Lin. No.:** 00018  
**Solution File No.:** 01421 **Soln. Date:** 01/12/2016 **Soln. No.:** 00242  
**Sequential File No.:** 01421 **File Date:** 01/12/2016

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDAB-0014  
**Control Solution %:** 0.100% **Expires:** 10/01/2017  
**Solution Control Lot:** 15J149 **Bottle No.:** 0961

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:00S		
Control 1 EC	0.101%	21:00S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:01S		
Control 2 EC	0.100%	21:02S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:02S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:02S		
Control 3 EC	0.099%	21:03S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:03S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: \_\_\_\_\_



Badge No.: 178

Date: 01/12/2016