

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00995	Calib. Date:	09/08/2015
Certification File No.:	00996	Cert. Date:	09/08/2015
Linearity File No.:	00997	Lin. Date:	09/08/2015
Solution File No.:	01085	Soln. Date:	01/17/2016
Sequential File No.:	01085	File Date:	01/17/2016
Calibrating Unit:	P	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	15J149	Expires:	10/01/2017
		Bottle No.:	0967

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:32S		
Control 1 EC	0.101%	02:33S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:33S		
Control 2 EC	0.099%	02:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:35S		
Control 3 EC	0.099%	02:36S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 01/17/2016