

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01363 **Calib. Date:** 09/08/2015 **Calib. No.:** 00026
Certification File No.: 01364 **Cert. Date:** 09/08/2015 **Cert. No.:** 00018
Linearity File No.: 01365 **Lin. Date:** 09/08/2015 **Lin. No.:** 00018
Solution File No.: 01429 **Soln. Date:** 01/20/2016 **Soln. No.:** 00245
Sequential File No.: 01429 **File Date:** 01/20/2016

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 10/01/2017
Solution Control Lot: 15J149 **Bottle No.:** 0974

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:26S		
Control 1 EC	0.099%	06:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	06:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:27S		
Control 2 EC	0.098%	06:28S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:28S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:28S		
Control 3 EC	0.097%	06:29S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:29S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:30S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: 

Badge No.: 174

Date: 01/20/2016