

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00995	Calib. Date:	09/08/2015
Certification File No.:	00996	Cert. Date:	09/08/2015
Linearity File No.:	00997	Lin. Date:	09/08/2015
Solution File No.:	01089	Soln. Date:	01/20/2016
Sequential File No.:	01089	File Date:	01/20/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	15J149	Expires:	10/01/2017
		Bottle No.:	0966

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:23S		
Control 1 EC	0.100%	05:24S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	05:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:25S		
Control 2 EC	0.100%	05:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:26S		
Control 3 EC	0.099%	05:27S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:28S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: _____



Badge No.: 174

Date: 01/20/2016