

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01363	Calib. Date:	09/08/2015
Certification File No.:	01364	Cert. Date:	09/08/2015
Linearity File No.:	01365	Lin. Date:	09/08/2015
Solution File No.:	01430	Soln. Date:	01/21/2016
Sequential File No.:	01430	File Date:	01/21/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	14C128	Expires:	03/26/2016
		Bottle No.:	0975

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:23S		
Control 1 EC	0.101%	13:23S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:24S		
Control 2 EC	0.099%	13:25S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:25S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:26S		
Control 3 EC	0.098%	13:26S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KLIMIK

First Name: DAVID

MI: W

Signature: *Tom H. D. Klimik* 77040

Badge No.: 7040

Date: 01/21/2016