

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01363 **Calib. Date:** 09/08/2015 **Calib. No.:** 00026
Certification File No.: 01364 **Cert. Date:** 09/08/2015 **Cert. No.:** 00018
Linearity File No.: 01365 **Lin. Date:** 09/08/2015 **Lin. No.:** 00018
Solution File No.: 01437 **Soln. Date:** 02/08/2016 **Soln. No.:** 00247
Sequential File No.: 01437 **File Date:** 02/08/2016

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 12/04/2016
Solution Control Lot: 14L134 **Bottle No.:** 1170

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:30S		
Control 1 EC	0.100%	04:31S	34.1 °C	*** TEST PASSED ***
Control 1 IR	0.100%	04:31S	34.1 °C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:32S		
Control 2 EC	0.098%	04:32S	34.1 °C	*** TEST PASSED ***
Control 2 IR	0.100%	04:32S	34.1 °C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:33S		
Control 3 EC	0.098%	04:34S	34.1 °C	*** TEST PASSED ***
Control 3 IR	0.099%	04:34S	34.1 °C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: _____



Badge No.: 178

Date: 02/08/2016