

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00995 Calib. Date: 09/08/2015 Calib. No.: 00019
Certification File No.: 00996 Cert. Date: 09/08/2015 Cert. No.: 00016
Linearity File No.: 00997 Lin. Date: 09/08/2015 Lin. No.: 00016
Solution File No.: 01105 Soln. Date: 02/20/2016 Soln. No.: 00209
Sequential File No.: 01105 File Date: 02/20/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 01/05/2017
Solution Control Lot: 16A150 Bottle No.: 0512

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:27S		
Control 1 EC	0.101%	22:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	22:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:28S		
Control 2 EC	0.100%	22:29S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:29S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:30S		
Control 3 EC	0.101%	22:30S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:30S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: 

Badge No.: 178

Date: 02/20/2016