

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01109	Calib. Date: 03/01/2016
Certification File No.:	01110	Calib. No.: 00020
Linearity File No.:	01111	Cert. Date: 03/01/2016
Solution File No.:	01126	Cert. No.: 00017
Sequential File No.:	01126	Lin. Date: 03/01/2016
		Lin. No.: 00017
		Soln. Date: 04/01/2016
		Soln. No.: 00212
		File Date: 04/01/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	16A150	Expires: 01/05/2017
		Bottle No.: 0883

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:15S		
Control 1 EC	0.102%	22:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:16S		
Control 2 EC	0.100%	22:17S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:18S		
Control 3 EC	0.100%	22:19S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	22:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:19S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHALLORAN

First Name: SEAN

MI: P

Signature: _____

Handwritten signature: S. O'Halloran 146

Badge No.: 146

Date: 04/01/2016