

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01109	Calib. Date:	03/01/2016
Certification File No.:	01110	Cert. Date:	03/01/2016
Linearity File No.:	01111	Lin. Date:	03/01/2016
Solution File No.:	01128	Soln. Date:	04/23/2016
Sequential File No.:	01128	File Date:	04/23/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXB S3-114
Control Solution %:	0.100%			Expires:	12/04/2016
Solution Control Lot:	14L134			Bottle No.:	1154

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:17D		
Control 1 EC	0.102%	06:17D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:18D		
Control 2 EC	0.101%	06:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:20D		
Control 3 EC	0.100%	06:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: _____



177

Badge No.: 177

Date: 04/23/2016