

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01109	Calib. Date: 03/01/2016
Certification File No.:	01110	Calib. No.: 00020
Linearity File No.:	01111	Cert. Date: 03/01/2016
Solution File No.:	01131	Cert. No.: 00017
Sequential File No.:	01131	Lin. Date: 03/01/2016
		Lin. No.: 00017
		Soln. Date: 04/30/2016
		Soln. No.: 00214
		File Date: 04/30/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	201603L	Expires: 03/22/2018
		Bottle No.: 1176

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:12D		
Control 1 EC	0.102%	00:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	00:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:13D		
Control 2 EC	0.102%	00:14D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.103%	00:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:15D		
Control 3 EC	0.102%	00:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	00:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: \_\_\_\_\_

Badge No.: 162

Date: 04/30/2016