

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01448 Calib. Date: 03/01/2016 Calib. No.: 00027
Certification File No.: 01449 Cert. Date: 03/01/2016 Cert. No.: 00019
Linearity File No.: 01450 Lin. Date: 03/01/2016 Lin. No.: 00019
Solution File No.: 01485 Soln. Date: 05/07/2016 Soln. No.: 00253
Sequential File No.: 01485 File Date: 05/07/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 03/22/2018
Solution Control Lot: 201603L Bottle No.: 1164

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 04:31D | | |
| Control 1 EC | 0.100% | 04:31D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 04:31D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:32D | | |
| Control 2 EC | 0.099% | 04:33D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 04:33D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:34D | | |
| Control 3 EC | 0.099% | 04:34D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 04:34D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:35D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: 

Badge No.: 177

Date: 05/07/2016