

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01109 Calib. Date: 03/01/2016 Calib. No.: 00020
Certification File No.: 01110 Cert. Date: 03/01/2016 Cert. No.: 00017
Linearity File No.: 01111 Lin. Date: 03/01/2016 Lin. No.: 00017
Solution File No.: 01138 Soln. Date: 05/11/2016 Soln. No.: 00215
Sequential File No.: 01138 File Date: 05/11/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 03/22/2018
Solution Control Lot: 201603L Bottle No.: 1174

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 07:48D | | |
| Control 1 EC | 0.103% | 07:49D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.103% | 07:49D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:49D | | |
| Control 2 EC | 0.101% | 07:50D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.102% | 07:50D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:51D | | |
| Control 3 EC | 0.102% | 07:51D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.102% | 07:51D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:52D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By: _____ MI: R
 Last Name: GRAMLICH First Name: JONATHAN
 Signature:  #179 Badge No.: 179
 Date: 05/11/2016