

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01109	Calib. Date:	03/01/2016
Certification File No.:	01110	Cert. Date:	03/01/2016
Linearity File No.:	01111	Lin. Date:	03/01/2016
Solution File No.:	01143	Soln. Date:	05/24/2016
Sequential File No.:	01143	File Date:	05/24/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	14L135	Expires:	12/31/2016
		Bottle No.:	0787

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:04D		
Control 1 EC	-.---%	18:05D	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	18:05D	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: \_\_\_\_\_



Badge No.: 165

Date: 05/24/2016