

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01448 Calib. Date: 03/01/2016 Calib. No.: 00027
Certification File No.: 01449 Cert. Date: 03/01/2016 Cert. No.: 00019
Linearity File No.: 01450 Lin. Date: 03/01/2016 Lin. No.: 00019
Solution File No.: 01498 Soln. Date: 05/25/2016 Soln. No.: 00254
Sequential File No.: 01498 File Date: 05/25/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 01/05/2017
Solution Control Lot: 16A150 Bottle No.: 1274

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:38D		
Control 1 EC	0.099%	09:38D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:39D		
Control 2 EC	0.099%	09:40D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		
Control 3 EC	0.097%	09:41D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A.

Signature: _____

PK #129

Badge No.: 129

Date: 05/25/2016