

# Calibrating Unit New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
Location: WALL TOWNSHIP POLICE  
Serial No.: ARXA-0037  
Calibration File No.: 01448 Calib. Date: 03/01/2016 Calib. No.: 00027  
Certification File No.: 01449 Cert. Date: 03/01/2016 Cert. No.: 00019  
Linearity File No.: 01450 Lin. Date: 03/01/2016 Lin. No.: 00019  
Solution File No.: 01501 Soln. Date: 05/28/2016 Soln. No.: 00255  
Sequential File No.: 01501 File Date: 05/28/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113  
Control Solution %: 0.100% Expires: 03/22/2018  
Solution Control Lot: 201603L Bottle No.: 1165

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:21D		
Control 1 EC	0.101%	00:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	00:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:22D		
Control 2 EC	0.101%	00:22D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	00:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:23D		
Control 3 EC	0.100%	00:24D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	00:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: 

Badge No.: 178

Date: 05/28/2016