

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01109	Calib. Date:	03/01/2016
Certification File No.:	01110	Cert. Date:	03/01/2016
Linearity File No.:	01111	Lin. Date:	03/01/2016
Solution File No.:	01152	Soln. Date:	06/02/2016
Sequential File No.:	01152	File Date:	06/02/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	201603L	Expires:	03/22/2018
		Bottle No.:	1177

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:53D		
Control 1 EC	0.101%	14:53D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:54D		
Control 2 EC	0.101%	14:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	14:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55D		
Control 3 EC	0.101%	14:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	14:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: _____



Badge No.: 174

Date: 06/02/2016