

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01448	Calib. Date: 03/01/2016
Certification File No.:	01449	Calib. No.: 00027
Linearity File No.:	01450	Cert. Date: 03/01/2016
Solution File No.:	01512	Cert. No.: 00019
Sequential File No.:	01512	Lin. Date: 03/01/2016
		Lin. No.: 00019
		Soln. Date: 06/11/2016
		Soln. No.: 00257
		File Date: 06/11/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	14L135	Expires: 12/31/2016
		Bottle No.: 0179

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:53D		
Control 1 EC	0.101%	04:54D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:55D		
Control 2 EC	0.100%	04:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	04:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:56D		
Control 3 EC	0.100%	04:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	04:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: _____



Badge No.: 177

Date: 06/11/2016