

# Calibrating Unit New Standard Solution Report

Serial No.: ARZL-0146

**Equipment** Alcotest 7110 MKIII-C

Location: WALL TOWNSHIP POLICE

Calibration File No.: 01109

Certification File No.: 01110

Linearity File No.: 01111

Solution File No.: 01160

Sequential File No.: 01160

Calib. No.: 00020

Cert. No.: 00017

Lin. No.: 00017

Soln. No.: 00219

File Date: 06/15/2016

Calib. Date: 03/01/2016

Cert. Date: 03/01/2016

Lin. Date: 03/01/2016

Soln. Date: 06/15/2016

File Date: 06/15/2016

Serial No.: DDXB S3-114

Expires: 03/22/2018

Bottle No.: 1172

Model No.: CU-34

Calibrating Unit: WET

Control Solution %: 0.100%

Solution Control Lot: 201603L

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:39D	33.9°C	*** TEST PASSED ***
Control 1 EC	0.101%	02:39D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.103%	02:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:40D	33.9°C	*** TEST PASSED ***
Control 2 EC	0.101%	02:41D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	02:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:41D	33.9°C	*** TEST PASSED ***
Control 3 EC	0.101%	02:42D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	02:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:43D	33.9°C	*** TEST PASSED ***

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: 

Badge No.: 177

Date: 06/15/2016